

**Anti-SARS-CoV-2 ELISA (IgG)****Instruction for use****For in vitro diagnostic use** IVD

ORDER NO.	ANTIBODIES AGAINST	IG CLASS	SUBSTRATE	FORMAT
EI 2606-9601 G	SARS-coronavirus-2 (SARS-CoV-2)	IgG	Ag-coated microplate wells	96 x 01 (96)

**Intended use**

The EUROIMMUN Anti-SARS-CoV-2 ELISA (IgG) is an enzyme-linked immunosorbent assay intended for the qualitative in vitro determination of human antibodies of the immunoglobulin class IgG against SARS-CoV-2 in human serum and plasma (EDTA, heparin or citrate) in the general population.

- The assay is for use in conjunction with the testing strategy outlined by the respective public health authorities in charge.
- Negative results do not exclude SARS-CoV-2 infection and should not be used as the sole basis for patient management decisions.
- False positive results for IgG antibodies may occur due to cross-reactivity from pre-existing antibodies or other possible causes.
- The test may also detect a response to vaccination against SARS-CoV-2.
- This assay is not intended to be used for screening patients or as an aid for diagnosis of patients with suspected COVID-19 infection.
- This assay is not intended for home-testing (or self-testing).
- Negative results must be combined with clinical observations, patient history and epidemiological information.
- False negative results can occur in elderly and immunocompromised patients.

Clinical significance

Severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2, previously called 2019-nCoV) belongs to the family of coronaviruses and, like SARS-CoV, is classified in the genus Betacoronavirus [1]. The new coronavirus originated in China in the city of Wuhan, Hubei province. It caused an infection wave, which has spread rapidly within the country and worldwide [2, 3]. Just a few days after the first report about patients with pneumonia of unclear origin, the causative pathogen was identified as SARS-CoV-2 [2-5].

SARS-CoV-2 is predominantly transmitted by droplet infection via coughing or sneezing and through close contact with infected persons [2-4, 6]. Health care personnel and family members are especially at risk of infection [6, 7]. The zoonotic reservoir of the virus appears to be bats [2, 4, 6].

The incubation time of SARS-CoV is three to seven, maximally 14 days [2]. The symptoms of SARS-CoV-2 infection are fever, coughing, breathing difficulties and fatigue [2-4, 6]. In most patients the infection manifests with symptoms of a mild febrile illness with irregular lung infiltrates. Some patients, especially elderly or chronically ill patients, develop acute respiratory distress syndrome (ARDS) [2, 3, 5, 6]. The fatality rate is between 0.6% and 7.2%, depending on the country [5]. In February 2020, the disease caused by SARS-CoV-2 was named COVID-19 by the WHO.

EI_2606G_A_CA_C04.docx

Version: 2021-09-22



Suitable methods for the diagnosis of SARS-CoV-2 infections are detection of viral RNA by reverse transcriptase polymerase chain reaction (RT-PCR) or of viral protein by ELISA primarily in sample material from the upper (nasopharyngeal or oropharyngeal smear) or lower respiratory tract (bronchoalveolar lavage fluid, tracheal secretion, sputum, etc.). The determination of antibodies enables confirmation of SARS-CoV-2 infection in patients with typical symptoms and in suspected cases. It also contributes to monitoring and outbreak control [4, 5]. For significant serological results, two patient samples should be investigated, one from the acute phase (week 1 of the illness) and one from the convalescent phase (3 to 4 weeks later) [4, 8, 9].

Cross-reactions with antibodies within the genus Betacoronavirus have been described [4, 5]. Currently, there is no medication or vaccine available against infection with this new virus [2, 7].

Antigen

The reagent wells of the ELISA were coated with an S1 domain of the spike protein of SARS-CoV-2 expressed recombinantly in the human cell line HEK 293.

Test principle

The test kit contains microplate strips each with 8 break-off reagent wells coated with recombinant structural protein of SARS-CoV-2. In the first reaction step, diluted patient samples are incubated in the wells. In the case of positive samples, specific IgG (also IgA and IgM) antibodies will bind to the antigens. To detect the bound antibodies, a second incubation is carried out using an enzyme-labelled anti-human IgG (enzyme conjugate) catalysing a colour reaction.

Contents of the test kit

Component	Colour	Format	Symbol
1. Microplate wells coated with antigens 12 microplate strips each containing 8 individual break-off wells in a frame, ready for use	-	12 x 8	
2. Calibrator (IgG, human), ready for use	dark red	1 x 2.0 ml	
3. Positive control (IgG, human), ready for use	blue	1 x 2.0 ml	
4. Negative control (IgG, human), ready for use	green	1 x 2.0 ml	
5. Enzyme conjugate peroxidase-labelled anti-human IgG, ready for use	green	1 x 12 ml	
6. Sample buffer , 0.09% sodium azide ready for use	light blue	1 x 100 ml	
7. Wash buffer 10x concentrate	colourless	1 x 100 ml	
8. Chromogen/substrate solution TMB/H ₂ O ₂ , ready for use	colourless	1 x 12 ml	
9. Stop solution 0.5 M sulphuric acid, ready for use	colourless	1 x 12 ml	
10. Protective foil	-	3 pieces	
11. Quality control certificate	-	1 protocol	-
12. Test instruction	-	1 booklet	-





Literature

1. Coronaviridae Study Group of the International Committee on Taxonomy of Viruses. **The species Severe acute respiratory syndrome-related coronavirus: classifying 2019-nCoV and naming it SARS-CoV-2.** Nat Microbiol. 2020; 5(4): 536-44
2. Wang G, Jin X. **The progress of 2019 Novel Coronavirus (2019-nCoV) event in China.** J Med Virol. 2020; 92(5): 468-72
3. Gralinski LE, Menachery VD. **Return of the Coronavirus: 2019-nCoV.** Viruses 2020, 12(2), 135
4. Udugama B, Kadhiresan P, Kozłowski HN, Malekjahani A, Osborne M, Li VYC, et al. **Diagnosing COVID-19: The Disease and Tools for Detection.** ACS Nano. 2020 Apr 9.
5. Cheng MP, Papenburg J, Desjardins M, Kanjilal S, Quach C, Libman M, et al. **Diagnostic Testing for Severe Acute Respiratory Syndrome-Related Coronavirus-2: A Narrative Review.** Ann Intern Med. 2020 Apr 13
6. Xiao SY, Wu Y, Liu H. **Evolving status of the 2019 novel coronavirus infection: proposal of conventional serologic assays for disease diagnosis and infection monitoring.** J Med Virol. 2020; 1-4
7. WHO: **Clinical management of severe acute respiratory infection when novel coronavirus (2019-nCoV) infection is suspected. Interim guidance,** 28 January 2020
8. WHO: **Laboratory testing for 2019 novel coronavirus (2019-nCoV) in suspected human cases. Interim guidance,** 17 January 2020
9. Okba NMA, Müller MA, Li W, Wang C, Geurts van Kessel CH, Corman VM, et al. **Severe Acute Respiratory Syndrome Coronavirus 2-Specific Antibody Responses in Coronavirus Disease 2019 Patients.** Emerg Infect Dis. 2020; 26(7)
10. Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DKW, et al. **Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR.** Euro Surveill. 2020; 25(3): pii=2000045

Technical support

In case of technical problems you can obtain assistance via the EUROIMMUN website (www.euroimmun.com/contact).

Meaning of the symbols

Symbol	Meaning	Symbol	Meaning
	Microplate strips		Lot description
	Calibrator		Protect from sunlight
	Positive control		Storage temperature
	Negative control		Unopened usable until (YYYY-MM-DD)
	Conjugate		CE-labelled
	Sample buffer		Manufacturing date (YYYY-MM-DD)
	Wash buffer, 10x concentrate		Manufacturer
	Substrate		Observe instructions for use
	Stop solution		Order number
	Protective foil		Contents suffice for <n> analyses
	In vitro diagnostic medical device		Biological risks



Additional materials and equipment (not supplied in the test kit)

- Automatic microplate washer: recommended. Washing of the microplates can also be carried out manually.
- Microplate reader: wavelength of 450 nm, reference wavelength range from 620 nm to 650 nm
- Calibrated pipettes
- Pipette tips
- Stepper pipette: recommended for the pipetting of enzyme conjugate, substrate, and stop solution
- Distilled or deionised water
- Incubator: for incubation of the microplate at +37°C
- Incubator or water bath: recommended to warm the wash buffer
- Stop watch

Storage and stability

The test kit has to be stored at a temperature between +2°C and +8°C, do not freeze. Unopened, all test kit components are stable until the indicated expiry date.

In use stability following the first opening

After opening, the reagents are stable until the indicated expiry date when stored at +2°C to +8°C and protected from contamination, unless stated otherwise below.

Warnings and precautions

- The product must only be used by trained laboratory personnel in a clinical or research laboratory.
- If the packed reagents are visibly damaged, do not use the test kit.
- Before using the product, read the instruction for use carefully. Use only the valid version provided with the product.
- Do not substitute or mix the EUROIMMUN reagents with reagents from other manufacturers.
- Observe Good Laboratory Practice (GLP) and safety guidelines. Some of the reagents contain preservatives in non-declarable concentrations. Avoid eye and skin contact with samples and reagents. In case of eye or skin contact, rinse thoroughly with water. Remove and wash contaminated clothing. In case of ingestion, obtain medical advice.
- The calibrator and controls of human origin have tested negative for HBsAg, anti-HCV, anti-HIV-1 and anti-HIV-2. Nonetheless, all reagents should be treated as being a potential infection hazard and should be handled with care.

Preparation and stability of the samples

- **Samples:** Human serum or EDTA, heparin or citrate plasma.
- **Sample preparation: Patient samples** are diluted 1:101 in sample buffer.

For example: dilute 10 µl serum in 1.0 ml sample buffer and mix well by vortexing (sample pipettes are not suitable for mixing).



- **Stability of the patient samples:**

- stored at +2°C to +8°C: up to 14 days
- incubate diluted samples within one working day

Preparation and stability of the reagents

Note: All reagents must be brought to room temperature (+18°C to +25°C) approx. 30 minutes before use.

The thermostatically adjustable ELISA incubator must be set to +37°C ± 1°C.

- **Coated wells:** Ready for use. Tear open the resealable protective wrapping of the microplate at the recesses above the grip seam. Do not open until the microplate has reached room temperature to prevent the strips from moistening. Immediately replace the remaining wells of a partly used microplate in the protective wrapping and tightly seal with the integrated grip seam (Do not remove the desiccant bag).
Once the protective wrapping has been opened for the first time, the wells coated with antigens can be stored in a dry place and at a temperature between +2°C and +8°C for 4 months.
- **Calibrator and controls:** Ready for use. Mix reagents thoroughly before use.
- **Enzyme conjugate:** Ready for use. Mix the reagent thoroughly before use.
- **Sample buffer:** Ready for use.
- **Wash buffer:** The wash buffer is a 10x concentrate. If crystallisation occurs in the concentrated buffer, warm it to +37°C and mix well before dilution. The quantity required should be removed from the bottle using a clean pipette and diluted 1:10 with deionised or distilled water (1 part reagent plus 9 parts water).
For example: For 1 microplate strip, 5 ml concentrate plus 45 ml water.
The working-strength wash buffer is stable for 4 weeks if stored at +2°C to +8°C and handled properly.
- **Chromogen/substrate solution:** Ready for use. Close the bottle immediately after use, as the contents are sensitive to light. The chromogen/substrate solution must be clear on use. Do not use the solution if it is blue coloured.
- **Stop solution:** Ready to use.

Waste disposal

Patient samples, calibrators, controls and incubated microplate strips should be handled as infectious waste. All reagents must be disposed of in accordance with local disposal regulations.

Quality control

For every group of tests performed, the extinction readings of the calibrator and ratios determined for the positive and negative controls must lie within the limits stated for the relevant test kit lot. A quality control certificate containing these reference values is included. If the values specified for the controls are not achieved, the test results may be inaccurate and the test should be repeated.

Reference material

As no quantified international reference serum exists for antibodies against SARS-CoV-2, the calibration is performed in ratios which are a relative measure for the concentration of antibodies in serum or plasma.



Limitations of the procedure

- For a medical diagnosis, the serological test result should always be interpreted together with the clinical symptoms of the patient and other results, e.g. those of the direct pathogen detection. A negative serological test result does not exclude the presence of the disease.
- The pipetting volumes, incubation times, temperatures, and preparation steps given in the instruction for use must be adhered to.
- Correct performance of sample collection and storage is crucial for the test results.
- The test system is validated for the determination of anti-SARS-CoV-2 IgG in human serum or plasma only.
- The binding activity of the antibodies and the activity of the enzyme used are temperature-dependent. It is therefore recommended using a thermostatically adjusted ELISA incubator in all incubation steps. The higher the room temperature during the incubation steps, the greater will be the extinction. The same variations also apply to the incubation times. However, the calibrators are subject to the same influences, with the result that such variations will be largely compensated in the calculation of the result.
- Insufficient washing (e.g. less than 3 wash cycles, too small wash buffer volumes, or too short residence times) can lead to false high extinction readings.
- Residual liquid (>10 µl) in the reagent wells after washing can interfere with the substrate and lead to false low extinction readings.
- The partial or complete adjustment of the test system to the use of instruments for automated sample processing or other liquid handling devices may result in differences between the results obtained with automated processing and those obtained with manual procedure. It is the responsibility of the user to validate the instruments used so that they yield test result within the reliable range.
- Results are for the detection of SARS-CoV-2 antibodies. IgG antibodies to SARS-CoV-2 become detectable later following infection. At this time, it is unknown how long IgG antibodies may persist following infection.
- Positive results for IgG could occur after infection and can be indicative of acute or recent infection (and successful immune response to a vaccine, once developed).
- The presence of specific antibodies is a sign of previous or current infection.
- Laboratories are required to report all positive results to the appropriate public health authorities.
- Potential exogenous interferents such as antibiotics, antiviral, drug metabolites, and common over-the-counter drugs have not been tested for interference with the assay results.
- The performance of this device has not been assessed in a population vaccinated against COVID-19.
- This test identifies antibodies to the spike protein of the SARS-CoV-2 virus and is therefore unable to distinguish between previously infected individuals and vaccinated individuals.
- The performance of the device has not been assessed on specimens from individuals who have been infected with emerging new variants of SARS-CoV-2, including the UK SARS-CoV-2 variant, SARS-CoV-2 VOC 202012/01 (B.1.1.7) or the new South Africa SARS-CoV-2 variant, 501Y.V2.
- Cross reactivity testing with Rhinovirus and hMPV has not been completed.



Clinical performance

Diagnostic sensitivity: The sensitivity was determined by investigating 166 samples from 152 European patients, using the Anti-SARS-CoV-2 ELISA (IgG). In these patients, infections with SARS-CoV-2 had been confirmed by RT-PCR test [10] based on a sample taken at the early phase of infection. In samples taken prior to day 10 (time point after onset of symptoms or positive direct detection), the Anti-SARS-CoV-2 ELISA (IgG) showed a sensitivity of 43.7%. The sensitivity of the Anti-SARS-CoV-2 ELISA (IgG) in samples collected after 10 was 94.4%. Borderline results (n = 7) were not considered in the calculation.

Days after symptom onset or positive direct detection	EUROIMMUN Anti-SARS-CoV-2 ELISA (IgG)		
	Positive	Negative	Sensitivity
≤ 10	38	49	43.7%
> 10	68	4	94.4%

Evaluation	borderline = positive		borderline = negative		Without borderline	
	Value	95% CI	Value	95% CI	Value	95% CI
Specificity	99.1%	98.4% - 99.5%	99.6%	99.1% - 99.9%	99.6%	99.1% - 99.9%
Sensitivity	94.7%	86.9% - 98.5%	90.7%	81.7% - 96.2%	94.4%	86.4% - 98.5%
n	1419		1419		1409	

Specificity: The specificity of the Anti-SARS-CoV-2 ELISA (IgG) was determined by analysing 222 patient samples that were positive, for instance, for antibodies against other human pathogenic coronaviruses, other pathogens or for rheumatoid factors. Additionally, 1122 samples from blood donors, children and pregnant women obtained before the first occurrence of SARS-CoV-2 were analysed. In these panels, no SARS-CoV-2-specific antibodies should be detectable. The results in the borderline range (n = 7) were not included in the calculation, resulting in a specificity of the Anti-SARS-CoV-2 ELISA (IgG) of 99.6%.

Panel	n	EUROIMMUN Anti-SARS-CoV-2 ELISA IgG
		Specificity
Antibodies against influenza (recently vaccinated incl. progressions)	40	100.0%
Acute EBV infection & heterophilic antibodies	22	100.0%
Infections with other human pathogenic coronaviruses	23	100.0%
Rheumatoid factors	40	100.0%
Blood donors (Germany, 2010)	150	98.7%
Blood donors (Germany, 2017)	250	99.2%
Blood donors (China, 2013)	49	100.0%
Blood donors (US, 2017)	400	99.3%
Pregnant women (China, 2013)	99	99.0%
Pregnant women (Germany, 2019)	100	100.0%
Children	74	100.0%
Elderly people	97	100.0%
Total	1344	99.6%



Assay procedure

(Partly) manual test performance

Sample incubation: (1st step)

Transfer **100 µl** of the **calibrator, positive and negative controls or diluted patient samples** into the individual microplate wells according to the pipetting protocol. Incubate for **60 minutes** at **+37 °C ± 1 °C**.

For manual processing of microplate wells, cover the finished test plate with the protective foil. When using an automated microplate processor for incubation, follow the recommendations of the instrument manufacturer.

Washing:

Manual: Remove the protective foil. Empty the wells and subsequently wash **3 times using 300 µl of working-strength wash buffer** for each wash.

Automatic: Remove the protective foil. Wash the reagent wells **3 times with 450 µl of working-strength wash buffer** (program setting: e.g. TECAN Columbus Washer "Overflow Mode").

Leave the wash buffer in each well for 30 to 60 seconds per washing cycle, then empty the wells. After washing (manual and automated tests), thoroughly dispose of all liquid from the microplate by tapping it on absorbent paper with the openings facing downwards to remove all residual wash buffer.

Note:

Free positions on the microplate strip should be filled with blank wells of the same plate format as that of the parameter to be investigated.

Conjugate incubation: (2nd step)

Pipette **100 µl of enzyme conjugate** (peroxidase-labelled anti-human IgG) into each of the microplate wells. For manual test performance cover the reagent wells with the protective foil.

Incubate **30 minutes** at **+37°C ± 1°C**.

Washing:

Remove the protective foil. Empty the wells. Wash as described above.

Substrate incubation: (3rd step)

Pipette **100 µl of chromogen/substrate solution** into each of the microplate wells. Incubate for **30 minutes** at room temperature (+18°C to +25°C) protected from direct sunlight.

Stopping:

Pipette **100 µl of stop solution** into each of the microplate wells in the same order and at the same speed as the chromogen/substrate solution was introduced.

Measurement:

Photometric measurement of the colour intensity should be made at a **wavelength of 450 nm** and a reference wavelength between 620 nm and 650 nm **within 30 minutes of adding the stop solution** in order to obtain the **extinction-** for each well. Prior to measuring, slightly shake the microplate to ensure a homogeneous distribution of the solution.

Test performance using fully automated analysis devices

Sample dilution and test performance are carried out fully automatically using an analysis device. The incubation conditions programmed in the respective software authorised by EUROIMMUN may deviate slightly from the specifications given in the ELISA test instruction. However, these conditions were validated in respect of the combination of the EUROIMMUN Analyzer I and this EUROIMMUN ELISA. Validation documents are available on enquiry.

Note: Processing on other fully automated systems is possible but must be validated by the user.



Pipetting protocol

	1	2	3	4	5	6	7	8	9	10	11	12
A	C	P 6	P 14	P 22								
B	pos.	P 7	P 15	P 23								
C	neg.	P 8	P 16	P 24								
D	P 1	P 9	P 17									
E	P 2	P 10	P 18									
F	P 3	P 11	P 19									
G	P 4	P 12	P 20									
H	P 5	P 13	P 21									

The pipetting protocol for microplate strips 1 to 4 is an example for the **ratio-based analysis** of 24 patient sera (P 1 to P 24).

The calibrator (C), the positive (pos.) and negative (neg.) controls, and the patient samples have each been incubated in one well. The reliability of the ELISA test can be improved by duplicate determinations for each sample.

The wells can be broken off individually from the strips. This makes it possible to adjust the number of test substrates used to the number of samples to be examined and minimises reagent wastage.

Both positive and negative controls serve as internal controls for the reliability of the test procedure. They should be assayed with each test run.

Test evaluation

The extinction of the calibrator defines the upper limit of the reference range of non-infected persons (**cut-off**) recommended by EUROIMMUN. Values above the indicated cut-off are to be considered as positive, those below as negative.

Ratio-based analysis: Results can be evaluated by calculating a ratio of the extinction of the control or patient sample over the extinction of the calibrator. Calculate the ratio according to the following formula:

$$\frac{\text{Extinction of the control or patient sample}}{\text{Extinction of calibrator}} = \text{Ratio}$$

EUROIMMUN recommends interpreting results as follows:

Ratio <0,8: **negative**
Ratio ≥ 0.8 to <1.1: **borderline**
Ratio ≥1.1: **positive**

For duplicate determinations the mean of the two values should be taken. If the two values deviate substantially from one another, EUROIMMUN recommends retesting the samples.



Analytical performance

Measurement range:

Limit of blank (LoB): ratio 0.13

Limit of detection (LoD): ratio 0.14

LoB and LoD were defined based on EUROIMMUN current SOP for emergency use authorization.

Precision: Studies on the intra-lab precision were carried out according to CLSI guideline EP05-A3. Four samples (reactivity distributed over the entire measurement range) were measured. The precision is given as standard deviation (SD) and coefficient of variation (CV).

Intra-lab precision

	Sample 1		Sample 2		Sample 3		Sample 4	
Mean	Ratio 0.07		Ratio 1.12		Ratio 2.36		Ratio 5.20	
	SD	%CV	SD	%CV	SD	%CV	SD	%CV
<i>Repeatability</i>	0.012	16.0	0.060	5.4	0.091	3.9	0.231	4.4
<i>Between run</i>	0.000	0.0	0.021	1.9	0.058	2.4	0.168	3.2
<i>Within day</i>	0.012	16.0	0.063	5.7	0.108	4.6	0.285	5.5
<i>Between day</i>	0.002	2.3	0.060	5.4	0.174	7.4	0.089	1.7
<i>Within lab</i>	0.012	16.2	0.087	7.8	0.205	8.7	0.299	5.7

Cross-reactivity (analytical specificity): However, due to their close relationship cross-reactions between SARS-CoV(-1) and SARS-CoV-2 are likely. Sera from patients with SARS-CoV(-1), MERS-CoV, HCoV-229E, HCoV-NL63, HCoV-HKU1 or HCoV-OC43 infections were investigated to examine this further. As expected, there were pronounced cross-reactivities, especially with anti-SARS-CoV(-1) IgG antibodies. Cross-reactions to other human pathogenic coronaviruses were not observed.

Group	N (positive) / n (negative)	Cross-reactivity [%]
RSV	0/30	0
Adenovirus	1/29	3.3
Enterovirus	1/29	3.3

Interference: Hemolytic, lipemic and icteric samples showed no influence on the result up to concentrations of 10 mg/ml hemoglobin, 20 mg/ml triglycerides and 0.4 mg/ml bilirubin in this ELISA. The following endogenous potential interferents have not been tested for interference: hematocrit, human anti-mouse antibodies (HAMA) & antibodies developed against protein expression systems used to generate recombinant antigens. Potential exogenous interferents such as medications most often prescribed in the target patient population (antibiotics, antiviral and drug metabolites), and common over-the-counter drugs have not been tested for interference.