



DiaSorin S.p.A. - Via Crescentino snc - 13040 Saluggia (VC) - Italy
DiaSorin Inc. - Stillwater, Minnesota 55082-0285, U.S.A
www.diasorin.com
Tel. +39.0161.4871

LIAISON® SARS-CoV-2 Ag (REF 311500)

For *in vitro* diagnostic use only.

1. INTENDED USE

The LIAISON® SARS-CoV-2 Ag assay uses chemiluminescence immunoassay (CLIA) technology for the qualitative determination of SARS-CoV-2 nucleocapsid protein antigen in nasopharyngeal swab (NPS) collected in COPAN Universal Transport Media (UTM) and direct anterior nasal swab (NS) without transport media, in individuals suspected to have COVID-19 by their healthcare provider within the first ten days from the onset of symptoms.

LIAISON® SARS-CoV-2 Ag assay does not differentiate between SARS-CoV and SARS-CoV-2 or MERS-CoV.

Results are for the identification of SARS-CoV-2 nucleocapsid protein antigen. Antigen is generally detectable in upper respiratory specimens during the acute phase of infection. Positive results indicate the presence of viral antigens, but clinical correlation with patient history and other diagnostic information is necessary to determine infection status. Positive results do not rule out bacterial infection or co-infection with other viruses. The agent detected may not be the definite cause of disease.

Negative results should be treated as presumptive, and do not rule out SARS-CoV-2 infection and should not be used as the sole basis for treatment or patient management decisions, including infection control decisions. Negative results should be considered in the context of a patient's recent exposures, history, and the presence of clinical signs and symptoms consistent with COVID-19, and confirmed with a molecular assay, if necessary, for patient management.

The test has to be performed on the LIAISON® XL Analyzer only.

The LIAISON® SARS-CoV-2 Ag is intended for use by trained clinical laboratory personal specifically instructed and trained *in vitro* diagnostic procedures.

2. SUMMARY AND EXPLANATION OF THE TEST

Coronavirus disease (COVID-19) is an infectious disease caused by a newly discovered coronavirus.

The causative virus of the COVID-19 is called Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2). It is a new strain of coronavirus that has not been previously identified in humans. It spreads primarily through contact with an infected person through respiratory droplets generated when a person coughs or sneezes, or through droplets of saliva or discharge from the nose. Infection with SARS-CoV-2 can cause mild symptoms including a runny nose, sore throat, cough, and fever. However, it can be more severe for some people and can lead to pneumonia or breathing difficulties. The elderly, and people with pre-existing medical conditions (such as, diabetes and heart disease), appear to be more vulnerable to becoming severely ill with the virus.

The incubation period for COVID-19 is thought to range from 2-14 days following exposure, with most cases showing symptoms approximately 4-5 days after exposure⁽¹⁾.

This test identifies the presence of the SARS-CoV-2 in the specimen through the detection of nucleocapsid protein antigen. This antigen is generally detectable in upper respiratory specimens during the acute phase of infection.

3. PRINCIPLE OF THE PROCEDURE

The method for qualitative determination of SARS-CoV-2 Ag in specimens collected and processed through the indicated pre-analytical procedure, is a direct two-step sandwich chemiluminescence immunoassay (CLIA). Specific rabbit polyclonal antibodies to nucleocapsid antigen are used for coating magnetic particles (solid phase) and linked to an isoluminol derivative (isoluminol-antibody conjugate). During the first incubation, SARS-CoV-2 viral antigen present in calibrators, samples or controls binds to the conjugate. During the second incubation, the solid phase reacts with the SARS-CoV-2 viral antigen already bound to the conjugate. After the second incubation, the unbound material is removed with a wash cycle. Subsequently, the starter reagents are added, and a flash chemiluminescence reaction is thus induced. The light signal, and hence the amount of isoluminol-antibody conjugate, is measured by a photomultiplier in relative light units (RLU) and indicates the presence or absence of the SARS-CoV-2 Ag in calibrators, samples or controls.

4. MATERIALS PROVIDED

Reagent integral

Magnetic particles (2.5 mL)	SORB	Magnetic particles coated with rabbit polyclonal to SARS-CoV-2 nucleocapsid antigen, BSA, phosphate buffer, < 0.1% sodium azide.
Calibrator 1 (1.8 mL)	CAL1	Antiseptic agent, Recombinant nucleoprotein (from <i>E.coli</i>), BSA, detergents.
Calibrator 2 (1.8 mL)	CAL2	Antiseptic agent, Recombinant nucleoprotein (from <i>E.coli</i>), BSA, detergents, an inert blue dye.
Specimen Diluent (19 mL)	DIL SPE	BSA, phosphate buffer, detergents, ProClin® 300, preservatives, an inert yellow dye.
Conjugate (13 mL)	CONJ	Rabbit polyclonal to SARS-CoV-2 nucleocapsid antigen conjugated to an isoluminol derivative, human serum, BSA, phosphate buffer, detergents, ProClin® 300, preservatives.
Number of tests		100

All reagents are supplied ready to use. The order of reagents reflects the layout of containers in the Reagent Integral. ProClin® is a trademark of the Dow Chemical Company (Dow) or an affiliated company of Dow.

Materials required but not provided

LIAISON® XL Cuvettes ([REF] X0016).
 LIAISON® XL Disposable Tips ([REF] X0015) or
 LIAISON® Disposable Tips ([REF] X0055).
 LIAISON® XL Starter Kit ([REF] 319200) or
 LIAISON® EASY Starter Kit ([REF] 319300).
 LIAISON® Wash/System Liquid ([REF] 319100).
 LIAISON® XL Waste Bags ([REF] X0025).

Additional required materials:

LIAISON® Control SARS-CoV-2 Ag ([REF] 311501).
 LIAISON® SARS-CoV-2 Sample Inactivation Buffer ([REF] 311502).

Materials not supplied in the kit:

COPAN Diagnostics CLASSIQswab™ or COPAN Diagnostic FLOQswab™ nasal swab.
 COPAN Diagnostic FLOQswab™ Minitip nasopharyngeal swab transported in Universal Transport Media (UTM, COPAN Diagnostics).

5. WARNINGS AND PRECAUTIONS


- For *in vitro* diagnostic use.
- For professional use only.
- This test has been authorized only for the presence of SARS-CoV-2 Antigen not for any other viruses or pathogens.
- All human serum and plasma units used to produce the components provided in this kit have been tested for the presence of HBsAg, anti-HCV, anti-HIV-1, anti-HIV-2, and found to be non-reactive. However, as no test method can offer absolute assurance that pathogens are absent, all specimens of human origin should be considered potentially infectious and handled with care.
- Use appropriate precautions in the collection, handling and storage of patient samples. Refer to CDC Interim Guidelines for Collection, Handling and Transportation of clinical specimens from persons with Coronavirus Disease 2019 (COVID-19) at <https://www.cdc.gov/coronavirus/2019-nCoV/lab/guidelines-clinical-specimens.html>, and to WHO's Interim guidance for Laboratory testing for coronavirus disease (COVID-19) in suspected human cases at <http://www.who.int/publications-detail/laboratory-testing-for-2019-novel-coronavirus-in-suspected-human-cases-20200117>, as amended and supplemented. Refer to the WHO website for additional publications.
- Used swabs must be treated as infectious waste.
- All samples, **even after the pre-analytical inactivation procedure**, and reagents containing biological materials used for the assay must be considered as potentially able to transmit infectious agents; accordingly samples, reagents and the waste must be handled with utmost care and disposed of in compliance with the laboratory guidelines and the statutory provisions in force in each country. Any materials for reuse must be appropriately sterilized in compliance with the local laws and guidelines. Check the effectiveness of the sterilization/decontamination cycle.
- For sample handling please refer to Section 8 - Specimen Collection and Preparation
- Observe the normal precautions required for handling all laboratory reagents.
- Do not eat, drink, smoke or apply cosmetics during the assay.
- Do not pipette by mouth.
- Strict adherence to the LIAISON® SARS-CoV-2 Ag assay instructions is necessary to obtain accurate results.
- Avoid direct contact with potentially infectious substances by wearing appropriate personal protective equipment such as laboratory coats, goggles, and disposable gloves. Wash hands thoroughly after removal of gloves.
- Avoid splashing or aerosolization of samples or reagents as droplets are a means of transmission of SARS-CoV-2 virus. All drops and spills must be wiped up with an appropriate disinfectant such as a sodium hypochlorite solution with 0.5% active chlorine, and all soiled materials must be disposed of as infectious waste.
- Visually inspect the integral vials for leaks at the membrane seals or elsewhere. If the vials are found to be leaking, discard them and the local customer service should be notified immediately.

- All waste associated with biological samples, biological reagents and disposable materials used for the assay must be considered potentially infectious and therefore should be disposed of in accordance with the national, state or local regulations and guidelines of the agencies holding jurisdiction over the laboratory.
- The LIAISON® XL Analyzer should be cleaned and decontaminated on a routine basis. See the LIAISON® XL Analyzer Operator's Manual for the cleaning and decontamination procedures.
- Previously frozen test samples, once thawed, must be thoroughly mixed prior to testing.
- Do not pool the contents of different vials of the same reagent (even if the reagents are from the same lot).
- Do not use kits or components beyond the expiration date indicated on the label.

Chemical Hazard and Safety Information

Reagents in this kit are classified in accordance with the US OSHA Hazard Communication Standard; individual US State Right-to-Know laws; Canadian Centre for Occupational Health and Safety Controlled Products Regulations; and European Union EC Regulation 1272/2008 (CLP) (for additional information see Safety Data Sheet available on www.diasorin.com).

Pursuant to EC Regulation 1272/2008 (CLP), hazardous reagents are classified and labeled as follows:

REAGENTS:	[DILSPE] , [CONJ]
CLASSIFICATION:	Skin sens. 1 H317 Aquatic chronic 3 H412
SIGNAL WORD:	Warning
SYMBOLS / PICTOGRAMS:	 GHS07 Exclamation mark
HAZARD STATEMENTS:	H317 May cause an allergic skin reaction. H412 Harmful to aquatic life with long lasting effects.
PRECAUTIONARY STATEMENTS:	P261 Avoid breathing dust/fume/gas/mist/vapours/spray. P280 Wear protective gloves/protective clothing/eye protection/face protection. P273 Avoid release to the environment. P362 Take off contaminated clothing and wash before reuse.
CONTAINS: (only substances prescribed pursuant to Article 18 of EC Regulation 1272/2008).	reaction mass of: 5-chloro-2-methyl-4-isothiazolin-3-one [EC no. 247-500-7] and 2-methyl-2H-isothiazol-3-one [EC no. 220-239-6] (3:1) (ProClin® 300).

Reagent containing sodium azide (Magnetic Particles **[SORB]**)

Sodium azide may react with lead or copper plumbing to form highly explosive metal azides. Immediately after disposal, flush with a large volume of water to prevent azide build-up. For further information, refer to "Decontamination of Laboratory Sink Drains to Remove Azide Salts", in the Manual Guide-Safety Management No. CDC-22 issued by the Centers for Disease Control and Preventions, Atlanta, GA, 1976.

Pursuant to EC Regulation 1272/2008 (CLP), **[SORB]** is labeled as EUH210, safety data sheets available on request. For additional information, see Safety Data Sheets available on www.diasorin.com.

6. REAGENT PREPARATION

REAGENT INTEGRAL

Please note the following important reagent handling precautions:

Resuspension of magnetic particles

Magnetic particles must be completely resuspended before the Reagent Integral is placed on the LIAISON® XL Analyzer. Follow the steps below to ensure complete resuspension:

- Before the seal is removed, rotate the small wheel at the magnetic particle vial compartment until the color of the suspension has changed to brown.
- Gentle and careful side-to-side mixing may assist in the resuspension of the magnetic particles (avoid foam formation).
- Visually check the bottom of the magnetic particle vial to confirm that all settled magnetic particles have been resuspended.
- Carefully wipe the surface of each septum to remove residual liquid.
- Repeat all steps as necessary until the magnetic particles are completely resuspended.

Foaming of Reagents

In order to ensure optimal performance of the Reagent Integral, foaming of all reagents should be avoided. Follow the steps below to prevent foaming of reagents:

- Visually inspect the reagents, calibrators in particular (located in position two and three following the magnetic particle vial), to ensure there is no foaming present before using the Reagent Integral.
- If foam is present after resuspension of the magnetic particles, place the integral on the LIAISON® XL Analyzer and allow the foam to dissipate.
- The Reagent Integral is ready for use once the foam of all reagents has dissipated and the integral is positioned onboard the LIAISON® XL Analyzer and mixing.

CONTROLS

Refer to the LIAISON® Control SARS-CoV-2 Ag Instructions for Use section for proper preparation and handling instructions.

7. REAGENT INTEGRAL STORAGE AND STABILITY

- **Sealed:** Stable at 2-8°C until the expiry date.
- **Opened on board or at 2-8°C:** stability up to 2 weeks.
- Use the storage rack provided with the analyzer for the upright storage of the Reagent Integral.
- Do not freeze.
- Keep upright for storage to facilitate later proper resuspension of the magnetic particles.
- Keep away from direct light.

8. SPECIMEN COLLECTION AND PREPARATION

Acceptable specimen types include:

- Dry Nasal swab (NS) processed following the below indicated pre-analytical procedure.
- Nasopharyngeal Swab (NPS) transported in Universal Transport Media (UTM), then processed following the pre-analytical procedure indicated below.

WARNING: for the collection and handling of swab specimens from the upper and lower respiratory tract, refer to the CDC Interim Guidelines for Collection, Handling and Transportation of clinical specimens from persons with Coronavirus Disease 2019 (COVID-19), and to the WHO's Interim guidance for Laboratory testing for coronavirus disease (COVID-19) in suspected human cases at <http://www.who.int/publications-detail/laboratory-testing-for-2019-novel-coronavirus-in-suspected-human-cases-20200117>, as amended and supplemented. Refer to the WHO website for additional publications.

PRE-ANALYTICAL PROCEDURE WORKFLOW FOR DRY SWAB

1. Collect the sample by nasal swab and place it dry into its container.
2. Dry swab placed into the container can be stored for up to 6 hours at 2°-8°C, prior to transferring the sample into the inactivation buffer.
3. All the following pre-analytical steps should be performed at room temperature (15°-25°C) using the universal precautions for handling potentially infectious specimens.
4. All clinical samples and reagents must be at room temperature (15°-25°C) before beginning the next step of the procedure.
5. Place and soak the swab in the pre-filled tube containing the LIAISON® SARS-CoV-2 Sample Inactivation Buffer ([REF] 311502).
6. Roll the swab at least 5 times while pressing the head against the bottom and side of the tube.
7. Leave the swab in the tube with the Sample Inactivation Buffer for at least 1 minute.
8. Roll the head of the swab at least 5 times against the inside wall of the tube and remove it.
9. Dispose of the used swab in the biohazardous waste collection.
10. Cap the tube and incubate for at least 120 minutes before handling it.
Warning: Ensuring that the incubation time lasts at least 120 minutes is essential to properly manage the virus inactivation process and reduce the risk of potential infection. The Laboratory is responsible for the proper recording of this inactivation timing for each of the specimens. Samples that have undergone the pre-analytical inactivation process should be handled and disposed of as though they were potentially infectious.
11. Remove the cap and place the tube on board the instrument for testing. Samples should be tested as soon as possible after the completion of the pre-analytical procedure. If immediate testing is not possible, inactivated samples can be stored at 2°-8°C for up to 6 days, or at -20°C or colder for up to 1 month prior to testing. If samples are stored frozen, mix thawed samples well before testing. Frozen samples can undergo up to three freeze/thaw cycles without experiencing any change in performance. Samples after thawing may require centrifugation (i.e. 3,000g x 10') before testing.

PRE-ANALYTICAL PROCEDURE WORKFLOW FOR SWAB IN UNIVERSAL TRANSPORT MEDIUM (UTM)

1. Collect the sample by nasopharyngeal swab (not provided in the kit) transported in UTM.
2. Nasopharyngeal swab in UTM can be stored for up to 12 hours at 2°-8°C, prior to transferring the sample into the inactivation buffer.
3. All the clinical samples and reagents must be at room temperature (15°-25°C) before beginning the next step of the procedure.
4. All following pre-analytical steps should be performed at room temperature (15°-25°C) using universal precautions for handling potentially infectious specimens.
5. Add 1 mL of the specimen eluted in UTM into the tube containing the Sample Inactivation Buffer ([REF] 311502).
6. Cap the tube and mix the specimen by vortex for 5-10 sec.
7. Incubate the tube at RT for at least 120 minutes before handling it.
Warning: Ensuring that the incubation time lasts at least 120 minutes is essential to properly manage the virus inactivation process and reduce the risk of potential infection. The Laboratory is responsible for the proper recording of this inactivation timing for each of the specimens. Samples that have undergone the pre-analytical inactivation process should be handled and disposed of as though they were potentially infectious.
8. Remove the cap and place the tube on board the instrument for testing. Samples should be tested as soon as possible after the completion of the pre-analytical procedure. If immediate testing is not possible, inactivated samples can be stored at 2°-8°C for up to 5 days or at -20°C or colder for up to 1 month prior to testing. If samples are stored frozen, mix thawed samples well before testing. Frozen samples can undergo up to one freeze/thaw cycles without experiencing any change in performance. Freshly collected specimens or samples after thawing, may require centrifugation (i.e. 3,000 g x 10') before testing.

Inadequate sample collection, handling, storage or transport may yield erroneous results.

For both procedures, the minimum volume required for a single determination is 400 µL of specimen (100 µL specimen + 300 µL dead volume).

9. CALIBRATION

By testing the assay specific calibrator, the detected relative light unit (RLU) values can adjust the assigned master curve. Each calibration solution is sufficient for performing 4 calibrations.

Recalibration in triplicate is mandatory whenever at least one of the following conditions occurs:

- A new lot of Starter Kit is used.
- The previous calibration was performed more than 1 week before.
- Each time a new lot of integral is used.
- The analyzer has been serviced.
- Control values lie outside the expected ranges.

LIAISON® XL Analyzer: Calibrator values are stored in the Reagent Integral Radio Frequency IDentification transponder (RFID Tag).

10. ASSAY PROCEDURE

Strict adherence to the analyzer operator's manual ensures proper assay performance.

Each test parameter is identified via information encoded in the Reagent Integral Radio Frequency IDentification transponder (RFID Tag). In the event that the RFID Tag cannot be read by the analyzer, the integral cannot be used. Do not discard the Reagent Integral; contact your local DiaSorin technical support for instructions.

The analyzer operations are as follows:

1. Dispense the specimens (calibrator or control) and conjugate into the reaction cuvettes.
2. Incubate.
3. Dispense the Specimen Diluent and magnetic particles into the reaction cuvettes.
4. Incubate and wash.
5. Add the Starter Reagents and measure the light emitted.

11. QUALITY CONTROL

The LIAISON® Control SARS-CoV-2 Ag (REF 311501) is recommended for the determination of quality control requirements for this assay and should be run in singlicate to monitor the assay performance.

Quality control is recommended once per day of use, or in accordance with local, state, and/or federal regulations or accreditation requirements and your laboratory's quality control procedures. It is recommended the user refer to CLSI document C24-A3 and 42 CFR 493.1256(c) for guidance on appropriate quality control practices⁽⁷⁾.

The range of concentrations of each control is reported on the certificate of analysis and indicates the limits established by DiaSorin for control values that can be obtained in reliable assay runs.

The performance of other controls should be evaluated for compatibility with this assay before they are used. Appropriate value ranges should be established for all quality control materials used.

Quality control could be performed by running the LIAISON® Control SARS-CoV-2 Ag:

- at least once per day of use,
- whenever the kit is calibrated,
- whenever a new lot of Starter Reagents is used.

Control values must lie within the expected ranges: whenever one of the controls lies outside the expected ranges, calibration should be repeated and the controls retested. If control values obtained after successful calibration repeatedly lie outside the predefined ranges, the test should be repeated using an unopened control vial. If control values lie outside the expected ranges, patient results must not be reported.

12. INTERPRETATION OF RESULTS

The analyzer automatically calculates results expressed as AU up to 10⁵ AU and grades the results as negative or positive. For details, refer to the analyzer operator's manual.

Numerical results should not be reported to health care providers.

Sample results should be interpreted as follows:

LIAISON® SARS-CoV-2 Ag assay		
AU	Result	Rules and interpretation
< 200.00	Negative	A result below 200 AU may indicate the absence of SARS-CoV-2 antigen in the specimen.
≥ 200.00	Positive	A result above or equal to 200 AU generally indicates presence of the SARS-CoV-2 antigen in the specimen.

A Negative result does not rule out infection by SARS-CoV-2. Negative results should be considered in the context of patient's history, presence of clinical signs and symptoms consistent with COVID-19, other diagnostic procedures available to the physician, and potential recent exposures to infected subjects. Despite a Negative result, confirmation with a molecular assay may be performed if necessary for patient management.

A Negative result may occur if the sample was collected, stored or transported improperly.

A Positive result indicates the presence of viral antigens, but clinical correlation with the patient's history and other diagnostic information is necessary to determine the infection status.

A Positive result does not rule out bacterial infection or co-infection with other viruses.

The agent detected may not be the definite cause of disease.

A failure to follow the indicated pre-analytical procedures may adversely affect the test performance.

13. LIMITATIONS

1. For prescription use only.
2. For *in vitro* diagnostic use.
3. The performance of this device has not been assessed in a population vaccinated against COVID-19.
4. This test detects both viable (live) and non-viable, SARS-CoV, and SARS-CoV-2. Test performance depends on the amount of virus (antigen) in the sample and may or may not correlate with viral culture results performed on the same sample.
5. Performance has not been established for use with specimens other than nasal swabs or nasopharyngeal swabs. Other specimen types have not been evaluated and should not be used with this assay.
6. A negative test result may occur if the level of antigen in a sample is below the detection limit of the test or if the sample was collected or transported improperly.
7. If the differentiation of specific SARS viruses and strains is needed, additional testing, in consultation with state or local public health departments, is required.
8. A skillful technique and strict adherence to the instructions are necessary to obtain reliable results.
9. Bacterial contamination or heat inactivation of the specimens may affect the test results.
10. The results obtained with this test should only be interpreted in conjunction with clinical findings, and the results from other laboratory tests and evaluations. This is especially important if the patient has had recent exposure to COVID-19, or clinical presentation indicates that COVID-19 is likely and diagnostic tests for other causes of illness (e.g., other respiratory illness) are negative. In this case, direct testing for the SARS-CoV-2 virus (e.g. PCR testing) should be considered.
11. Positive results do not rule out co-infections with other pathogens.
12. Positive test results do not differentiate between SARS-CoV-2 and SARS-CoV.
13. The performance of the device has not been assessed on specimens from individuals who have been infected with emerging variants of SARS-CoV-2 of public health concern.

14. SPECIFIC PERFORMANCE CHARACTERISTICS

14.1. Analytical specificity

Analytical specificity may be defined as the ability of the assay to accurately detect specific analyte in the presence of potentially interfering factors in the sample matrix (e.g. anticoagulants, haemolysis, effects of sample treatment), or cross-reactants.

Endogenous and Exogenous substances Interference

A study was performed to demonstrate that potentially endogenous and exogenous interfering substances potentially found in the upper respiratory tract do not cross-react or interfere in the LIAISON® SARS-CoV-2 Ag assay, at the indicated concentrations.

Interfering substance	Active Ingredient	Concentration
Whole Blood	Blood	4%
Mucin	Mucin Protein	0.5%
Sore Throat Spray	Chloraseptic (Menthol/Benzocaine)	1.5 mg/mL
Nasal Gel	Naso GEL (NeilMed)	5% v/v
Nasal Spray 1	CVS Nasal Drops (Phenylephrine)	15% v/v
Nasal Spray 2	Afrin (Oxymetazoline)	15% v/v
Nasal Spray 3	CVS Nasal Spray (Cromolyn)	15% v/v
Homeopathic Cold Remedy	Zicam	5% v/v
Nasal wash	Homeopathic (Alkalol)	1:10 dilution
Anti-Bacterial system	Tobramycin	4 µg/mL
Antibacterial	Mupirocin	7.5 mg/mL
Nasal Spray 4	Fluticasone Propionate	5% v/v
Anti viral drug	Tamiflu (Oseltamivir Phosphate)	5 mg/mL
Sore Throat Phenol Spray	Chloraseptic (Phenol)	15% v/v

Cross-reactivity and Interference by microorganisms and viruses

A cross-reactivity and potential interference study for the LIAISON® SARS-CoV-2 Ag assay was evaluated by testing various microorganisms and viruses with the LIAISON® SARS-CoV-2 Ag assay. Each organism and virus was tested in triplicate in the absence or presence of inactivated SARS-CoV-2. The tested concentration of the microorganisms and viruses are documented in the Table below.

Virus / Microorganism	Concentration tested	Cross-reactive result (Contains Virus/Micro-organism under evaluation)	Interference result (Contains Virus/Micro-organism under evaluation and inactivated SARS-CoV-2)
Adenovirus	10 ⁵ PFU/mL	Negative	Positive
Coronavirus 229E	10 ⁵ PFU/mL	Negative	Positive
Coronavirus NL63	10 ⁴ PFU/mL	Negative	Positive
Coronavirus OC43	10 ⁴ PFU/mL	Negative	Positive
Enterovirus 68	10 ⁵ PFU/mL	Negative	Positive
HMPV	10 ⁵ PFU/mL	Negative	Positive
Influenza A H1N1	10 ⁵ PFU/mL	Negative	Positive
Influenza A H3N2	10 ⁵ PFU/mL	Negative	Positive
Influenza B	10 ⁴ PFU/mL	Negative	Positive
MERS Coronavirus	0.0595 mg/mL	Positive	Positive
Parainfluenza Virus Type 1	10 ⁵ PFU/mL	Negative	Positive
Parainfluenza Virus Type 2	10 ⁵ PFU/mL	Negative	Positive
Parainfluenza Virus Type 3	10 ⁵ PFU/mL	Negative	Positive
Parainfluenza Virus Type 4b	10 ⁴ PFU/mL	Negative	Positive
Respiratory Syncytial Virus	10 ⁵ PFU/mL	Negative	Positive
Rhinovirus A	10 ⁵ PFU/mL	Negative	Positive
SARS Coronavirus	N/A*	Negative	Positive
<i>Bordetella pertussis</i>	10 ⁶ CFU/mL	Negative	Positive
<i>Candida albicans</i>	10 ⁶ CFU/mL	Negative	Positive
<i>Chlamydia pneumoniae</i>	10 ⁶ CFU/mL	Negative	Positive
<i>Haemophilus influenzae</i>	N/A*	Negative	Positive
<i>Legionella pneumophila</i>	10 ⁶ CFU/mL	Negative	Positive
<i>Mycobacterium tuberculosis</i>	10 ⁴ CFU/mL	Negative	Positive
<i>Mycoplasma pneumoniae</i>	10 ⁵ CFU/mL	Negative	Positive
<i>Pneumocystis Carinii</i>	5 ¹⁰ 6 nuclei/mL	Negative	Positive
<i>Staphylococcus aureus</i>	10 ⁶ CFU/mL	Negative	Positive
<i>Staphylococcus epidermidis</i>	10 ⁶ CFU/mL	Negative	Positive
<i>Streptococcus pneumoniae</i>	10 ⁴ CFU/mL	Negative	Positive
<i>Streptococcus pyogenes</i>	10 ⁴ CFU/mL	Negative	Positive
Pooled Human Nasal Wash	N/A*	Negative	Positive

* Material was not quantified; therefore no tested concentration can be reported. The material was tested at a dilution of 1:20. Human coronavirus HKU1 and Pneumocystis jirovecii were not available for wet testing. In-silico analysis (protein blast) was performed instead, and it was determined that no cross reactivity for these organisms is expected.

Based on the data generated during this study, only MERS Coronavirus demonstrated evidence of cross-reactivity with the LIAISON® SARS-CoV-2 Ag assay.

14.2. Precision

A five-day precision study was performed using a coded panel of 6 contrived samples to obtain samples across the assay range. Kit Controls were also included in the study. The panel of samples and kit controls were tested with the LIAISON® SARS-CoV-2 Ag assay in 6 replicates per run, 3 runs per day for 5 operating days on 2 kit lots, at 2 different sites, on 1 LIAISON® XL Analyzer at each site. The CLSI document EP5-A3 was consulted when preparing the testing protocol.

LIAISON® SARS-CoV-2 Ag Assay, Lot 1, Site 1										
Sample ID	N	Mean AU	Within Run		Between Run		Between Day		Overall	
			SD	%CV	SD	%CV	SD	%CV	SD	%CV
Kit Neg	90	1435.5*	41.5	2.9%	29.444	2.1%	49.712	3.5%	71.114	5.0%
Kit Pos	90	856.33	13.1	1.5%	12.341	1.4%	29.114	3.4%	34.225	4.0%
COVAG-01-U1	90	146.37	4.1	2.8%	1.307	0.9%	14.134	9.7%	14.788	10.1%
COVAG-01-U2	90	309.26	5.9	1.9%	2.07	0.7%	17.603	5.7%	18.695	6.0%
COVAG-01-U3	90	461.12	8.7	1.9%	6.443	1.4%	24.367	5.3%	26.677	5.8%
COVAG-01-U4	90	1108.9	18.7	1.7%	8.572	0.8%	24.278	2.2%	31.828	2.9%
COVAG-01-U5	90	8731.6	185.3	2.1%	531.327	6.1%	0**	0.0%	562.704	6.4%
COVAG-01-U6	90	70693	1478.4	2.1%	933.583	1.3%	1948.331	2.8%	2617.856	3.7%

* Elaboration performed on RLU values

LIAISON® SARS-CoV-2 Ag Assay, Lot 2, Site 2										
Sample ID	N	Mean AU	Within Run		Between Run		Between Day		Overall	
			SD	%CV	SD	%CV	SD	%CV	SD	%CV
Kit Neg	90	1398.4*	38.0	2.7%	12.781	0.9%	32.698	2.3%	51.756	3.7%
Kit Pos	90	989.71	15.0	1.5%	0	0.0%	26.854	2.7%	30.749	3.1%
COVAG-01-U1	90	174.88	4.5	2.6%	13.258	7.6%	3.537	2.0%	14.454	8.3%
COVAG-01-U2	90	355.86	4.9	1.4%	17.925	5.0%	16.747	4.7%	25.015	7.0%
COVAG-01-U3	90	515.49	8.7	1.7%	56.407	10.9%	53.339	10.3%	78.123	15.2%
COVAG-01-U4	90	1285	8.7	0.7%	56.407	4.4%	53.339	4.2%	78.123	6.1%
COVAG-01-U5	90	8076	157.1	1.9%	644.529	8.0%	1331.778	16.5%	1487.866	18.4%
COVAG-01-U6	90	61348	1146.9	1.9%	562.09	0.9%	4713.655	7.7%	4883.626	8.0%

* Elaboration performed on RLU values

LIAISON® SARS-CoV-2 Ag Assay, Two Lots Combined												
Sample ID	N	Mean AU	Within Run		Between Run		Between Day		Between-Lot		Overall	
			SD	%CV	SD	%CV	SD	%CV	SD	%CV	SD	%CV
Kit Neg	180	1417*	39.8	2.8%	22.7	1.6%	42.1	3.0%	16.7	1.2%	64.4	4.5%
Kit Pos	180	923.02	14.1	1.5%	8.4	0.9%	28.0	3.0%	93.4	10.1%	98.9	10.7%
COVAG-01-U1	180	160.63	4.3	2.7%	9.4	5.9%	10.3	6.4%	19.5	12.1%	24.4	15.2%
COVAG-01-U2	180	332.56	5.4	1.6%	12.8	3.8%	17.2	5.2%	31.9	9.6%	38.8	11.7%
COVAG-01-U3	180	488.31	8.7	1.8%	40.1	8.2%	41.5	8.5%	32.0	6.6%	66.6	13.6%
COVAG-01-U4	180	1197	20.7	1.7%	17.0	1.4%	15.8	1.3%	124.3	10.4%	128.1	10.7%
COVAG-01-U5	180	8404	171.8	2.0%	590.6	7.0%	938.5	11.2%	122.3	1.5%	1128.7	13.4%
COVAG-01-U6	180	66020	1323.1	2.0%	770.6	1.2%	3606.6	5.5%	6403.1	9.7%	7506.8	11.4%

* Elaboration performed on RLU values

These precision performance data refer to the repeatability and reproducibility of the analytical assay, not considering the variability of the sample collection and of the pre-analytical procedure of the sample.

The results refer to the groups of samples investigated and are not guaranteed specifications, as differences may exist between laboratories and locations.

14.3. Hook effect

The LIAISON® SARS-CoV-2 Ag assay was tested with a high concentration sample and no high-dose hook effect was observed.

14.4. Limit of Detection (LoD)

The LoD was determined by evaluating different dilutions of heat inactivated SARS-CoV-2 virus added to pooled nasal wash.

Nasal swabs

50 µL of the viral particle solution was added to dry swabs and the swabs were then placed into 1 mL of inactivation buffer. The inactivation buffer with eluted viral particles was tested repeatedly using the LIAISON® SARS-CoV-2 Ag test (n=20).

Nasopharyngeal swab

50 µL of the viral particle solution was added to dry swab and the swab was then placed into 3 mL of the transport media (UTM). The transport media with eluted viral particles was then diluted 1:1 with inactivation buffer and tested repeatedly using the LIAISON® SARS-CoV-2 Ag test (n=20).

Limit of Detection (LoD) is defined as the lowest virus concentration at which a minimum of 19 replicates out of 20 generate a Positive result. Results are reported in the following table.

Swab type	LoD (TCID ₅₀ /mL)
Dry Nasal swab collected using CLASSIQswab™	300
Dry Nasal swab collected using FLOQswab™	300
Nasopharyngeal Swab collected using FLOQswab™ minitip	575

15. SUMMARY OF CLINICAL PERFORMANCE

Patient Demographics

Patient demographics (by age and gender) are available for the samples used in the studies.

The patient positivity breakdown based on age of the patient:

Age	LIAISON® SARS-CoV-2 Ag Nasal swabs (n= 141)			LIAISON® SARS-CoV-2 Ag Nasopharyngeal swabs (n= 185)		
	Total #	Total Positive	Prevalence	Total #	Total Positive	Prevalence
≤ 5 years	0	0	N/A	0	0	N/A
6 to 21 years	5	1	20.0%	11	3	27.3%
22 to 59 years	81	17	21.0%	125	39	31.2%
≥ 60 years	45	11	24.4%	43	6	14.0%
unknown	10	3	30.0%	6	2	33.3%

Nasal swabs

The clinical performance of the LIAISON® SARS-CoV-2 Ag test was established with a total of 141 nasal swab samples collected from individual subjects during the 2020 COVID-19 pandemic. Specimens were collected from 2 different vendors (in the US) and 1 clinical center (in Europe) from symptomatic patients suspected of COVID-19. Nasal swabs were collected and eluted in Sample Inactivation Buffer ([REF] 311502) and stored frozen until tested. Samples were thawed and tested with LIAISON® SARS-CoV-2 Ag accordingly to the assay procedure. All subjects were confirmed as positive (≤10 days from onset of symptoms) or negative by a reference extracted RT-PCR method, used as comparator method for the study.

		Reference extracted RT-PCR assay		
		POS	NEG	Total
LIAISON® SARS-CoV-2 Ag on nasal swabs	POS	32	0	32
	NEG	1	108	109
	Total	33	108	141

Positive Percentage Agreement (Sensitivity): 32/33 (97.0%, Wilson 95% CI: 84.7 – 99.5%).

Negative Percentage Agreement (Specificity): 108/108 (100%, Wilson 95% CI: 96.6 – 100%).

Nasopharyngeal swabs

The clinical performance of the LIAISON® SARS-CoV-2 Ag test was established with a total of 185 nasopharyngeal swab collected from individual subjects during the 2020 COVID-19 pandemic. Specimens were collected from 3 different vendors (2 in the US, 1 in Europe) and 1 clinical center (in Europe) from symptomatic patients suspected of COVID-19. Nasopharyngeal swabs were collected in transport media, stored frozen, thawed and eluted in Sample Inactivation Buffer (REF 311502), and tested with LIAISON® SARS-CoV-2 Ag accordingly to the assay procedure. All subjects were confirmed as positive (≤ 10 days from onset of symptoms) or negative by a reference extracted RT-PCR method, used as comparator method for the study.

		Reference extracted RT-PCR assay		
		POS	NEG	Total
LIAISON® SARS-CoV-2 Ag on nasopharyngeal swabs	POS	49	1	50
	NEG	2	133	135
	Total	51	134	185

Positive Percentage Agreement (Sensitivity): 49/51 (96.1%, Wilson 95% CI: 86.8 – 98.9%).

Negative Percentage Agreement (Specificity): 133/134 (99.5%, Wilson 95% CI: 95.9 – 99.9%).

The clinical performance of the LIAISON® SARS-CoV-2 Ag test was evaluated with positive results stratified by the RT-PCR method cycle threshold (Ct) counts. As presented in the table below, the positive agreement of the LIAISON® SARS-CoV-2 Ag test is higher with samples of a Ct count <33.

		Reference extracted RT-PCR assay		
		< 33Ct	≥ 33 Ct	Total
LIAISON® SARS-CoV-2 Ag on nasopharyngeal swabs	POS	45	4	49
	NEG	1	1	2
	Total	46	5	51

Positive Percentage Agreement (Sensitivity) for Ct <33: 45/46 (97.8%, Wilson 95% CI 88.7 – 99.6%).

LIAISON® Control SARS-CoV-2 Ag (REF 311501)

1. INTENDED USE

The LIAISON® SARS-CoV-2 Ag controls (negative and positive) are intended for use as assayed quality control samples to monitor the performance and reliability of the LIAISON® SARS-CoV-2 Ag assay. The performance characteristics of the LIAISON® SARS-CoV-2 Ag controls have not been established for any assays or instrument platforms other than LIAISON® XL.

The certificate of analysis barcodes give specific information on the lot of controls and should be read by the hand-held barcode scanner of the LIAISON® XL Analyzer prior to loading the control vials on board. For details, refer to the analyzer operator's manual.

For *in vitro* diagnostic use only.

2. MATERIALS PROVIDED

Negative control (2 x 2.7 mL)	CONTROL-	Two (2) vials. Antiseptic agent, BSA, detergents.
Positive control (2 x 2.7 mL)	CONTROL+	Two (2) vials. Antiseptic agent, recombinant nucleoprotein (from <i>E.coli</i>), BSA, detergents.

All reagents are supplied ready to use. The range of values of each control is reported on the certificate of analysis and indicates the limits established by DiaSorin for control values that can be obtained in reliable assay runs. Each laboratory is responsible for adopting different limits to meet individual requirements.

3. WARNINGS AND PRECAUTIONS

- For prescription use only.
- For *in vitro* diagnostic use.
- This test has not been FDA cleared or approved; this test has been authorized by FDA under an EUA for use by laboratories certified under CLIA that meet requirements to perform moderate or high complexity tests.
- This test has been authorized only for the presence of SARS-CoV-2 Antigen not for any other viruses or pathogens.
- Controls are not kit lot specific and may be safely interchanged even with different Reagent Integral lots.
- Observe the normal precautions required for handling all laboratory reagents.
- Dispose of all waste material in accordance with local guidelines.

4. SAFETY PRECAUTIONS

Do not eat, drink, smoke or apply cosmetics in the assay laboratory.

Do not pipette by mouth.

Avoid direct contact with potentially infectious material by wearing laboratory clothing, protective goggles and disposable gloves. Wash hands thoroughly at the end of each assay.

Avoid splashing or forming an aerosol. All drops of biological reagent must be removed with a sodium hypochlorite solution with 0.5% active chlorine, and the means used must be treated as infectious waste.

All samples and reagents containing biological materials used for the assay must be considered as potentially able to transmit infectious agents. The waste must be handled with care and disposed of in compliance with the laboratory guidelines and the statutory provisions in force in each country. Any materials for reuse must be appropriately sterilized in compliance with the local laws and guidelines. Check the effectiveness of the sterilization/decontamination cycle.

Do not use kits or components beyond the expiration date stated on the label.

5. STORAGE AND STABILITY

Upon receipt, the controls must be stored at 2-8°C in an upright position to prevent adherence of the solution to the vial cap. Do not freeze.

When the controls are stored sealed, and kept upright, they remain stable at 2-8°C until the expiry date.

Once opened, the controls are stable for up to 4 weeks when properly stored at 2-8°C between two successive uses.

Avoid bacterial contamination of the controls.

The controls should not be used past the expiry date indicated on the vial labels.

6. PREPARATION OF REAGENTS

- Place the control vials in type C racks on the analyzer. Each control vial is sufficient for performing at least 20 tests.
- The dead volume is 400 µL.
- At the time of use, equilibrate the controls to room temperature (20-25°C) before opening the vials and keep them on board the instrument only for the amount of time required for quality control testing.
- After use, stopper the vials promptly and store them at 2-8°C in an upright position.
- During handling, use appropriate precautions to avoid bacterial contamination of the controls.

7. LIMITATIONS

Control values for assays other than the LIAISON® SARS-CoV-2 Ag assay have not been established. If users wish to use this control material with other assays, it is their responsibility to establish the appropriate ranges.

The performance of other controls should be evaluated for compatibility with this assay before they are used. Appropriate reference ranges should be established for all quality control materials used.

If the control values obtained after successful calibration repeatedly lie outside the expect ranges, the test should be repeated using an unopened control vial.

8. HANDLING

For proper handling please refer to the analyzer operator's manual.

9. ASSIGNED VALUES

The ranges of SARS-CoV-2 Ag concentration in the controls are printed on the certificate of analysis. They have been established after taking into account run variability, in order to guarantee the accuracy of analytical results and to obtain indications on stability or deterioration of reagents.

LIAISON® SARS-CoV-2 Sample Inactivation Buffer (REF 311502)

1. INTENDED USE

Sample Inactivation Buffer is used in association with nasal swab (NS) and nasopharyngeal swab (NPS) (dry or eluted in UTM) and is intended to decrease the viral load of SARS-CoV-2 in the specimen by inactivation, during the pre-analytical procedure, when testing the samples with the LIAISON® SARS-CoV-2 Ag assay.

The performance characteristics of LIAISON® SARS-CoV-2 Sample Inactivation Buffer have not been established for any assays or instrument platforms other than LIAISON® XL.

For *in vitro* diagnostic use only.

2. MATERIALS PROVIDED

Sample Inactivation Buffer (100 x 1.0 mL)	BUF	Antiseptic agent, detergents and inert blue dyes.
--	------------	---

Supplied ready to use.

3. WARNINGS AND PRECAUTIONS

- For prescription use only.
- For use under an Emergency Use Authorization Only.
- For *in vitro* diagnostic use.
- This test has not been FDA cleared or approved; this test has been authorized by FDA under an EUA for use by laboratories certified under CLIA that meet requirements to perform moderate or high complexity tests.
- This test has been authorized only for the presence of SARS-CoV-2 Antigen not for any other viruses or pathogens.
- Sample shall be collected following the manufacturer's instructions for use and the COVID-19 standard of care.
- Observe the normal precautions required for handling all laboratory reagents.
- Dispose of all waste material in accordance with local guidelines.

4. SAFETY PRECAUTIONS

Do not eat, drink, smoke or apply cosmetics in the assay laboratory.

Do not pipette by mouth.


Avoid direct contact with potentially infected material by wearing laboratory clothing, protective goggles, and disposable gloves. Wash hands thoroughly at the end of each assay.

Avoid splashing or forming an aerosol. All drops of biological reagent must be removed with a sodium hypochlorite solution with 0.5% active chlorine, and the means used must be treated as infectious waste.

All samples and reagents containing the biological materials used for the assay must be considered as potentially able to transmit infectious agents. The waste must be handled with care and disposed of in compliance with the laboratory guidelines and the statutory provisions in force in each country. Any materials for reuse must be appropriately sterilized in compliance with the local laws and guidelines. Check the effectiveness of the sterilization/decontamination cycle.

Do not use reagents beyond the expiration date stated on the label.

5. STORAGE AND STABILITY

Keep away from sunlight. 

Upon receipt, the vials must be stored at 2-8°C in an upright position to prevent adherence of the solution to the vial cap.

Do not freeze.

When the vials are stored sealed and kept upright, they remain stable at 2°-8°C until the expiry date.

Vials can be stored for up to eight (8) weeks at 15-25°C.

Avoid bacterial contamination.

The reagent should not be used past the expiry date indicated on the vial labels.

6. LIMITATIONS

Do not use to inactivate any viruses other than SARS-CoV-2.

7. HANDLING

For proper handling please refer to the LIAISON® SARS-CoV-2 Ag Instructions for Use.

Manufactured for:

DiaSorin Inc.

1951 Northwestern Ave,

Stillwater, MN, 55082-0285 USA.

For Customer Service in Canada and the U.S. call toll free: 1-800-328-1482.

REFERENCES

1. Lauer SA, Grantz KH, Bi Q, et al. The Incubation Period of Coronavirus Disease 2019 (COVID-19) From Publicly Reported Confirmed Cases: Estimation and Application. *Ann Intern Med.* 2020;172(9):577-582. doi:10.7326/M20-0504.

200/008-043, 01 - 2022-04