



SARS-CoV-2 Antigen Rapid Test

Package Insert

REF L031-125W5 English

A rapid test for the qualitative detection of SARS-CoV-2 nucleocapsid antigens in nasal and nasopharyngeal swab specimens.

For professional in vitro diagnostic use only.

INTENDED USE

The SARS-CoV-2 Antigen Rapid Test is a lateral flow chromatographic immunoassay for the qualitative detection of the nucleocapsid protein antigen from SARS-CoV-2 in nasal and nasopharyngeal swab specimens.

Results are for the identification of SARS-CoV-2 nucleocapsid antigen. This antigen is generally detectable in upper respiratory samples during the acute phase of infection.

Negative results, from patients with symptom beyond seven days, should be treated as presumptive and confirmed with a molecular assay, if necessary, for patient management.

The SARS-CoV-2 Antigen Rapid Test is intended for use by trained clinical laboratory personnel and individuals trained in point of care settings.

SUMMARY

The novel coronaviruses belong to the beta genus. COVID-19 is an acute respiratory infectious disease. People are generally susceptible. Currently, the patients infected by the novel coronavirus are the main source of infection.

PRINCIPLE

The SARS-CoV-2 Antigen Rapid Test is a qualitative membrane based chromatographic immunoassay for the qualitative detection of the nucleocapsid protein antigen from SARS-CoV-2 in human nasal and nasopharyngeal swab specimens.

When specimens are processed and added to the test cassette, SARS-CoV-2 antigens, if present in the specimen, will react with the anti-SARS-CoV-2 antibody-coated particles, which have been pre-coated on the test strip.

To serve as a procedure control, a colored line will always appear in the control line region indicating that proper volume of specimen has been added and membrane wicking has occurred.

REAGENTS

The test cassette contains anti-SARS-CoV-2 antibodies.

PRECAUTIONS

- For professional in vitro diagnostic use only. Do not use after the expiration date.
Do not eat, drink, or smoke in the area where the specimens or kits are handled.
Do not use the test if the pouch is damaged.
Handle all specimens as if they contain infectious agents.

Table with 3 columns: Chemical Name/Concentration, Harms (GHS) code for each ingredient, Concentration. Row 1: Sodium Azide, Acute Tox. 2 * (H300); Aquatic Acute 1 (H400); Aquatic Chronic 1 (H410), 0.02%

STORAGE AND STABILITY

- The kit can be stored at temperatures between 2 - 30 °C.
The test is stable until the expiration date printed on the sealed pouch.
The test must remain in the sealed pouch until use.
DO NOT FREEZE.
Do not use after the expiration date.

MATERIALS

Materials Provided

- Test Cassettes
Disposable Swabs (Nasal Swabs)*
Extraction Buffer Tubes
Package Insert
* The Disposable Swabs are produced by another manufacturer.

Materials Required But Not Provided


- Personal Protective Equipment
Timer

SPECIMEN COLLECTION AND PREPARATION

- Testing should be performed immediately after specimen collection, or at most within one (1) hour after

specimen collection, if stored at room temperature (15-30°C).

How to collect an anterior nasal swab sample:




- Carefully insert one of the Disposable Nasal Swabs, provided with your kit, into one nostril. Using gentle rotation, push the swab less than 2.5 cm (1 inch) from the edge of the nostril.
Rotate the swab 5 times against the mucosa inside the nostril to ensure sufficient specimen collection.
Using the same swab, repeat the process in the other nostril to ensure that an adequate amount of sample is collected from both nasal cavities.
Withdraw the swab from the nasal cavity. The specimen is now ready for preparation using the extraction buffer tubes.

DIRECTIONS FOR USE

Allow the test and extraction buffer to reach room temperature (15-30 °C) prior to testing.

- Use an extraction buffer tube for each specimen to be tested and label each tube appropriately.
Remove the aluminum foil from the top of extraction buffer tube.
Insert the swab into the tube and swirl it for 30 seconds. Then rotate the swab at least 5 times while squeezing the sides of the tube. Take care to avoid splashing contents out of the tube.
Remove the swab while squeezing the sides of the tube to extract the liquid from the swab.
Attach the dropper tip firmly onto the extraction buffer tube containing the sample. Mix thoroughly by swirling or flicking the bottom of the tube.
Remove the test cassette from the foil pouch and use it as soon as possible.
Place the test cassette on a flat and clean surface.
Add the processed specimen to the sample well of the test cassette.
a. Invert the extraction buffer tube with the dropper tip pointing downwards and hold it vertically. Gently squeeze the tube, dispensing 4 drops of the processed specimen into the sample well.
Wait for the colored line(s) to appear. The result should be read at 15-30 minutes. Do not read the result after 30 minutes.



INTERPRETATION OF RESULTS

(Please refer to the illustration above)

Repeat testing is needed to improve test accuracy. Please follow the table below when interpreting test results.

Table with 5 columns: Status on First Day of Testing, First Result Day 1, Second Result Day 3, Third Result Day 5, Interpretation. Rows include With Symptoms and Without Symptoms scenarios.

Results should be considered in the context of an individual's recent exposures, history, and the presence of clinical signs and symptoms consistent with COVID-19.

NEGATIVE: Only one colored control line appears in the control region (C). No apparent colored line appears in the test line region (T). This means that no SARS-CoV-2 antigen was detected.

To increase the chance that the negative result for COVID-19 is accurate, you should:
Test again in 48 hours if the individual has symptoms on the first day of testing.
Test 2 more times at least 48 hours apart if the individual does not have symptoms on the first day of testing.
A negative test result indicates that the virus that causes COVID-19 was not detected in the sample. A negative result does not rule out COVID-19. There is a higher chance of false negative results with antigen tests

compared to laboratory-based tests such as PCR tests. If the test is negative but COVID-19-like symptoms, e.g., fever, cough, and/or shortness of breath continue, follow up testing for SARS-CoV-2 with a molecular test or testing for other respiratory disease should be considered.

All negative results should be treated as presumptive and confirmation with a molecular assay may be necessary if there is a high likelihood of SARS-CoV-2 infection, such as in an individual with a close contact with COVID-19 or with suspected exposure to COVID-19 or in communities with high prevalence of infection.

POSITIVE:* Two distinct colored lines appear. One line in the control line region (C) and the other line-in the test line region (T). This means that the presence of SARS-CoV-2 antigen was detected. Repeat testing does not need to be performed if the patient has a positive result at any time.

*NOTE: The intensity of the color in the test line (T) may vary depending on the level of the SARS-CoV-2 antigen present in the specimen. Therefore, any shade of color in the test line region (T) should be considered positive.

INVALID: Control line fails to appear. Insufficient specimen volume or incorrect operation are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test cassette.

QUALITY CONTROL

Internal procedural controls are included in the test. A colored line appearing in the control line region (C) is an internal procedural control. It confirms sufficient specimen volume and correct procedural technique.

Control swabs are not supplied with this kit; however, it is recommended that positive and negative controls should be tested as a good laboratory practice to ensure that the test cassette and that the test procedure performed correctly.

LIMITATIONS

- The SARS-CoV-2 Antigen Rapid Test is for in vitro diagnostic use only. The test should be used for the detection of SARS-CoV-2 antigens in nasal and nasopharyngeal swab specimens only.
Specimens should be tested as quickly as possible after specimen collection and at most within the hour following collection.
Use of viral transport media may result in decreased test sensitivity.
A false-negative test may result if the level of antigen in a sample is below the detection limit of the test or if the sample was collected incorrectly.
Test results should be correlated with other clinical data available to the physician.
A positive test result does not rule out co-infections with other pathogens.
A positive test result does not differentiate between SARS-CoV and SARS-CoV-2.
A negative test result is not intended to rule out other viral or bacterial infections.
The performance of the SARS-CoV-2 Antigen Rapid Test has not been assessed in a population vaccinated against COVID-19.
Laboratories may be required to report all positive results in accordance with any country-specific or public health authority requirements.
Use in conjunction with the testing strategy outlined by public health authorities in your area.
This test is not intended for home testing (or self-testing).
The performance has not been established with all circulating variants but is anticipated to be reflective of the prevalent variants in circulation at the time and location of the clinical evaluation.

PERFORMANCE CHARACTERISTICS

Clinical Sensitivity, Specificity and Accuracy

This clinical performance data reflects the accuracy of the test when testing once. This test was not clinically validated for serial testing. The serial testing recommendations are supported by the study conducted by the National Institutes of Health (NIH) and the University of Massachusetts Chan Medical School in collaboration with the US FDA.

Nasal Swab Specimens

The performance of SARS-CoV-2 Antigen Rapid Test was established with 605 nasal swabs collected from individual patients who were suspected of COVID-19. The results show that the relative sensitivity and the relative specificity are as follows:

Table with 4 columns: Method, RT-PCR (Nasopharyngeal Swab Specimens) Results (Negative, Positive), Total Results. Rows include SARS-CoV-2 Antigen Rapid Test (Nasal Swab Specimens) and Total Results.

Relative Sensitivity: 97.1% (93.1%-98.9%)* Accuracy: 98.8% (97.6%-99.5%)* Relative Specificity: 99.5% (98.2%-99.9%)* *95% Confidence Intervals

Stratification of the prospective positive samples post onset of symptoms between 0-3 days has a positive percent agreement (PPA) of 98.3% (n=60) and 4-7 days has a PPA of 96.0% (n=25).

Prospective positive samples with Ct value <=30 has a positive percent agreement (PPA) of 100% (n=73) and Ct value >30 has a positive percent agreement (PPA) of 81.0% (n=21).

Nasopharyngeal Swab Specimens

The performance of SARS-CoV-2 Antigen Rapid Test was established with 299 nasopharyngeal swabs collected from individual patients who were suspected of COVID-19. The results show that the relative

sensitivity and the relative specificity are as follows:

| Method | | RT-PCR (Nasopharyngeal Swab Specimens) | | Total Results |
|---|----------|--|----------|---------------|
| SARS-CoV-2 Antigen Rapid Test (Nasopharyngeal Swab Specimens) | Results | Negative | Positive | |
| | Negative | 175 | 3 | 178 |
| | Positive | 1 | 120 | 121 |
| Total Results | | 176 | 123 | 299 |

Relative Sensitivity: 97.6% (92.8% - 99.5%)*
Accuracy: 98.7% (96.5% - 99.6%)*
Relative Specificity: 99.4% (96.5% - 99.9%)*
*95% Confidence Intervals

Stratification of the prospective positive samples post onset of symptoms between 0-3 days has a positive percent agreement (PPA) of 100% (n=20) and 4-7 days has a PPA of 100% (n=24).

Prospective positive samples with Ct value ≤30 has a positive percent agreement (PPA) of 100% (n=39) and Ct value >30 has a positive percent agreement (PPA) of 88.9% (n=9).

Serial-testing clinical performance

A prospective clinical study was conducted between January 2021 and May 2022 as a component of the Rapid Acceleration of Diagnostics (RADx) initiative from the National Institutes of Health (NIH). A total of 7,361 individuals were enrolled via a decentralized clinical study design, with a broad geographical representation of the United States. Per inclusion criteria, all individuals were asymptomatic upon enrollment in the study and at least 14 days prior to it and did not have a SARS-CoV-2 infection in the three months prior to enrollment. Participants were assigned to one of three EUA authorized SARS-CoV-2 OTC rapid antigen tests to conduct serial testing (every 48 hours) for 15 days. If an antigen test was positive, the serial-antigen testing result is considered positive.

At each rapid antigen testing time point, study subjects also collected a nasal swab for comparator testing using a home collection kit (using a 15-minute normalization window between swabs). SARS-CoV-2 infection status was determined by a composite comparator method on the day of the first antigen test, using at least two highly sensitive EUA RT-PCRs. If results of the first two molecular test were discordant a third highly sensitive EUA RT-PCR test was performed, and the final test result was based upon the majority rule. Study participants reported symptom status throughout the study using the MyDataHelps app. Two-day serial antigen testing is defined as performing two antigen tests 36 – 48 hours apart. Three-day serial antigen testing is defined as performing three antigen tests over five days with at least 48 hours between each test.

Out of the 7,361 participants enrolled in the study, 5,609 were eligible for analysis. Among eligible participants, 154 tested positive for SARS-CoV-2 infection based on RT-PCR, of which 97 (62%) were asymptomatic on the first day of their infection, whereas 57 (39%) reported symptoms on the first day of infection. Pre-symptomatic subjects were included in the positive percent agreement (PPA) of asymptomatic individuals, if they were asymptomatic on the first day of antigen testing, regardless of whether they developed symptoms at any time after the first day of testing.

Performance of the antigen test with serial testing in individuals is described in below table:

Data establishing PPA of COVID-19 antigen serial testing compared to the molecular comparator single day testing throughout the course of infection with serial testing. Data is from all antigen tests in study combined.

| Days after First PCR Positive Test Result | Asymptomatic on First Day of Testing | | | Symptomatic on First Day of Testing | | |
|---|---|------------------|------------------|-------------------------------------|------------------|------------------|
| | Ag Positive/PCR Positive (Antigen Test Performance % PPA) | | | | | |
| | 1 Test | 2 Tests | 3 Tests | 1 Test | 2 Tests | 3 Tests |
| 0 | 9/97 (9.3%) | 35/89 (39.3%) | 44/78 (56.4%) | 34/57 (59.6%) | 47/51 (92.2%) | 44/47 (93.6%) |
| 2 | 17/34 (50.0%) | 23/34 (67.6%) | 25/32 (78.1%) | 58/62 (93.5%) | 59/60 (98.3%) | 43/43 (100%) |
| 4 | 16/21 (76.2%) | 15/20 (75.0%) | 13/15 (86.7%) | 55/58 (94.8%) | 53/54 (98.1%) | 39/40 (97.5%) |
| 6 | 20/28 (71.4%) | 21/27 (77.8%) | 16/18 (88.9%) | 27/34 (79.4%) | 26/33 (78.8%) | 22/27 (81.5%) |
| 8 | 13/23 (56.5%) | 13/22 (59.1%) | 4/11 (36.4%) | 12/17 (70.6%) | 12/17 (70.6%) | 7/11 (63.6%) |
| 10 | 5/9 (55.6%) | 5/8 (62.5%) | N/A | 4/9 (44.4%) | 3/7 (42.9%) | N/A |

1 Test= one (1) test performed on the noted days after first PCR positive test result. Day 0 is the first day of documented infection with SARS-CoV-2.

2 Tests= two (2) tests performed an average of 48 hours apart. The first test performed on the indicated day and the second test performed 48 hours later.

3 Tests= three (3) tests performance an average of 48 hours apart. The first test performed on the indicated day, the second test performed 48 hours later, and a final test performed 48 hours after the second test.

Limit of Detection (LOD)

The LOD of SARS-CoV-2 Antigen Rapid Test was established using limiting dilutions of an inactivated viral sample. The viral sample was spiked with negative human nasal and nasopharyngeal sample pool into a series of concentrations. Each level was tested for 30 replicates. The results show that the LOD is 1.6*10² TCID₅₀/mL.*

* Base on the concentration of virus in extraction buffer

Cross-Reactivity (Analytical Specificity) and Microbial Interference

Cross-reactivity was evaluated by testing a panel of related pathogens and microorganisms that are likely to be present in the nasal cavity. Each organism and virus were tested in the absence or presence of heat-inactivated SARS-CoV-2 virus at low positive level. Both nasal swab specimens and nasopharyngeal swab specimens were tested.

No cross-reactivity or interference was observed with the following microorganisms when tested at the concentration presented in the table below. The SARS-CoV-2 Antigen Rapid Test does not differentiate between SARS-CoV and SARS-CoV-2.

| Potential Cross-Reactant | Test Concentration | Cross-Reactivity (in the absence of SARS-CoV-2 virus) | Interference (in the presence of SARS-CoV-2 virus) |
|--------------------------|------------------------|---|--|
| Virus | Adenovirus | 1.14 x 10 ⁶ TCID ₅₀ /mL No 3/3 negative | No 3/3 positive |
| | Enterovirus | 9.50 x 10 ⁵ TCID ₅₀ /mL No 3/3 negative | No 3/3 positive |
| | Human coronavirus 229E | 1.04 x 10 ⁵ TCID ₅₀ /mL No 3/3 negative | No 3/3 positive |
| | Human coronavirus OC43 | 2.63 x 10 ⁵ TCID ₅₀ /mL No 3/3 negative | No 3/3 positive |
| | Human coronavirus NL63 | 1.0 x 10 ⁵ TCID ₅₀ /mL No 3/3 negative | No 3/3 positive |
| | Human | 1.25 x 10 ⁵ TCID ₅₀ /mL No | No |

| | | | | |
|--------------------------------------|-------------------------------|---|--------------------|--------------------|
| Bacteria | Metapneumovirus | | 3/3 negative | 3/3 positive |
| | MERS-coronavirus | 7.90 x 10 ⁵ TCID ₅₀ /mL | No 3/3 negative | No 3/3 positive |
| | Influenza A | 1.04 x 10 ⁵ TCID ₅₀ /mL | No 3/3 negative | No 3/3 positive |
| | Influenza B | 1.04 x 10 ⁵ TCID ₅₀ /mL | No 3/3 negative | No 3/3 positive |
| | Parainfluenza virus 1 | 1.25 x 10 ⁵ TCID ₅₀ /mL | No 3/3 negative | No 3/3 positive |
| | Parainfluenza virus 2 | 3.78 x 10 ⁵ TCID ₅₀ /mL | No 3/3 negative | No 3/3 positive |
| | Parainfluenza virus 3 | 1.0 x 10 ⁵ TCID ₅₀ /mL | No 3/3 negative | No 3/3 positive |
| | Parainfluenza virus 4 | 2.88 x 10 ⁶ TCID ₅₀ /mL | No 3/3 negative | No 3/3 positive |
| | Respiratory syncytial virus | 3.15 x 10 ⁵ TCID ₅₀ /mL | No 3/3 negative | No 3/3 positive |
| | Rhinovirus | 3.15 x 10 ⁵ TCID ₅₀ /mL | No 3/3 negative | No 3/3 positive |
| | Bordetella pertussis | 2.83 x 10 ⁹ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Chlamydia trachomatis | 3.13 x 10 ⁸ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Haemophilus influenza | 1.36 x 10 ⁸ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Legionella pneumophila | 4.08 x 10 ⁹ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Mycobacterium tuberculosis | 1.72 x 10 ⁷ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Mycoplasma pneumoniae | 7.90 x 10 ⁷ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Staphylococcus aureus | 1.38 x 10 ⁷ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Staphylococcus epidermidis | 2.32 x 10 ⁹ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Streptococcus pneumoniae | 1.04 x 10 ⁸ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Streptococcus pyogenes | 4.10 x 10 ⁶ CFU/mL | No 3/3 negative | No 3/3 positive |
| Pneumocystis jirovecii-S. cerevisiae | 8.63 x 10 ⁷ CFU/mL | No 3/3 negative | No 3/3 positive | |
| Pseudomonas aeruginosa | 1.87 x 10 ⁸ CFU/mL | No 3/3 negative | No 3/3 positive | |
| Chlamydia pneumoniae | 1×10 ⁶ IFU/ml | No 3/3 negative | No 3/3 positive | |
| Yeast | Candida albicans | 1.57 x 10 ⁸ CFU/mL | No 3/3 negative | No 3/3 positive |
| | Pooled human nasal wash | | No 3/3 negative | No 3/3 positive |

To estimate the likelihood of cross-reactivity with SARS-CoV-2 of organisms that were not available for wet testing, in-silico analysis was used to assess the degree of protein sequence homology. The comparison between SARS-CoV-2 nucleocapsid protein and human coronavirus HKU1 revealed a low homology of 36.7% across 82.8% of the SARS-CoV-2 nucleocapsid sequence. The result suggests that cross-reactivity with human coronavirus HKU1 cannot be completely ruled out.

Interfering Substances

The following substances, naturally present in respiratory specimens or that may be artificially introduced into the nasal cavity or nasopharynx, were evaluated. Each substance was tested in the absence or presence of SARS-CoV-2 virus at low positive level. The final concentration of the substances tested are listed below and were found not to affect test performance.

| Interfering Substance | Active Ingredient | Concentration | Results (in the absence of SARS-CoV-2 virus) | Results (in the presence of SARS-CoV-2 virus) |
|---|---|---------------|--|---|
| Endogenous | Biotin | 2.4 mg/mL | 3/3 negative | 3/3 positive |
| | Mucin | 0.5% w/v | 3/3 negative | 3/3 positive |
| | Whole Blood | 4% v/v | 3/3 negative | 3/3 positive |
| Afrin Original Nasal Spray | Oxymetazoline | 15% v/v | 3/3 negative | 3/3 positive |
| ALKALOL Allergy Relief Nasal Spray | Homeopathic | 1:10 Dilution | 3/3 negative | 3/3 positive |
| Chloraseptic Max Sore Throat Lozenges | Menthol, Benzocaine | 1.5 mg/mL | 3/3 negative | 3/3 positive |
| CVS Health Fluticasone Propionate Nasal Spray | Fluticasone propionate | 5% v/v | 3/3 negative | 3/3 positive |
| Equate Fast-Acting Nasal Spray | Phenylephrine | 15% v/v | 3/3 negative | 3/3 positive |
| Equate Sore Throat Phenol Oral Anesthetic Spray | Phenol | 15% v/v | 3/3 negative | 3/3 positive |
| Original Extra Strong Menthol Cough Lozenges | Menthol | 1.5 mg/mL | 3/3 negative | 3/3 positive |
| NasalCrom Nasal Spray | Cromolyn | 15% v/v | 3/3 negative | 3/3 positive |
| NeilMed NasoGel for Dry Noses | Sodium Hyaluronate | 5% v/v | 3/3 negative | 3/3 positive |
| Throat Lozenge | Dyclonine Hydrochloride | 1.5mg/mL | 3/3 negative | 3/3 positive |
| Zicam Cold Remedy | Galphimia glauca, Luffa operculata, Sabadilla | 5% v/v | 3/3 negative | 3/3 positive |
| Antibiotic | Mupirocin | 10 mg/mL | 3/3 negative | 3/3 positive |
| Tamiflu | Oseltamivir Phosphate | 5 mg/mL | 3/3 negative | 3/3 positive |

| | | | | |
|--------------------------------------|--------------------|---------|--------------|--------------|
| Antibiotic | Tobramycin | 4 µg/mL | 3/3 negative | 3/3 positive |
| Mometasone Furoate Nasal Spray | Mometasone Furoate | 5%v/v | 3/3 negative | 3/3 positive |
| Physiological Seawater Nasal Cleaner | NaCl | 15%v/v | 3/3 negative | 3/3 positive |

PRECISION

Intra-Assay
Within-run precision was determined using 60 replicates of specimens: negative control and SARS-CoV-2 antigen positive controls. The specimens were correctly identified >99% of the time.

Inter-Assay
Between-run precision was determined using 60 independent assays on the same specimen: negative specimen and SARS-CoV-2 antigen positive specimen. Three different lots of the SARS-CoV-2 Antigen Rapid Test were tested using these specimens. The specimens were correctly identified >99% of the time.

High Dose Hook Effect

No high dose hook effect was observed when tested with up to a concentration of 1.43 x 10⁵ TCID₅₀/mL of heat-inactivated SARS-CoV-2 virus with the SARS-CoV-2 Antigen Rapid Test.

POC STUDY

A total of 9 operators from 3 sites performed the SARS-CoV-2 Antigen Rapid Test on 60 blinded labeled specimens by following the instructions of the package insert and recorded the results on data sheet. Based on the results of this POC clinical evaluation, untrained operators with various background and experience levels can perform the SARS-COV-2 Antigen Rapid Test correctly after read the product package insert without other training. The untrained operators found that the test procedure described in the package insert is simple to follow.

BIBLIOGRAPHY

- Shuo Su, Gary Wong, Weifeng Shi, et al. Epidemiology, Genetic recombination, and pathogenesis of coronaviruses. Trends in Microbiology, June 2016, vol. 24, No. 6: 490-502
- Susan R. Weiss, Julian L. Leibowitz, Coronavirus Pathogenesis, Advances in Virus Research, Volume 81: 85-164

Index of Symbols

| | | | | | |
|--|------------------------------------|--|-----------------------------------|--|-------------------|
| | Manufacturer | | Contains sufficient for <n> tests | | Temperature limit |
| | In vitro diagnostic medical device | | Use-by date | | Do not reuse |
| | Consult instructions for use | | Batch code | | Catalogue number |
| | Date of manufacture | | | | |



ACON Biotech (Hangzhou) Co., Ltd.

No.210 Zhenzhong Road, West Lake District
Hangzhou, P.R. China, 310030